

Mechanical Alignment of a Customized TKA Implant System as Determined by Intraoperative Computer Navigation

Gary A. Levensgood, MD

INTRODUCTION

A neutral mechanical limb alignment has been linked to increased Total Knee Arthroplasty (TKA) survivorship^{1,2}. Patient-specific cutting blocks have demonstrated close to neutral alignment on average, but have also found a varying number of outliers ($>3^\circ$) within specific brands (Otismed, Visionaire, Signature etc.) as well as between brands³⁻⁶. However, no study has investigated patient-specific cutting guides in conjunction with customized TKA implants (ConforMIS iTotal CR). Hence, the purpose of this study was to utilize computer navigation intra-operatively to measure the mechanical alignment of a patient-specific implant and instrument TKA system.

METHODS

A consecutive series of 63 patients undergoing TKA, utilizing patient-specific instruments and implants, were prospectively measured with intraoperative computer navigation. The instruments and implants were first created utilizing a pre-operative CT scan. The mechanical alignment of all patients was measured using computer navigation. The patient-specific instruments were then utilized per the manufacturer's recommendations and bone cuts were made (Figure 1). All bone cuts were recorded utilizing the navigation system as a confirmatory measurement. The patient-specific implants were then fixated and surgery completed. Final mechanical alignment was recorded with computer navigation.



Figure 1. Computer assisted navigation was used to measure the pre-op and post-op mechanical alignment intra-operatively.

RESULTS

The patient-specific instruments and implants provided neutral mechanical alignment, zero degrees, in 84.1% of patients (53/63). The remaining 10 patients each had a post-operative alignment within $\pm 2^\circ$ of neutral, with no

outliers. The average pre-operative degree of deformity for this cohort was 5.57° v. 0.19° postoperatively ($p < 0.0001$) (Table 1). The mean correction angle for this cohort was 5.70° . Additionally, no patients had extension deficits as measured with navigation post-operatively (7.5° pre-op for 40/63 patients).

Table 1: Average pre-op and post-op mechanical axis measurements

	Pre-op	Post-op
Average Deformity	$5.57^\circ \pm 6.5^\circ$	$0.19^\circ \pm 0.47^\circ$
Deformity Range	12° varus - 15° valgus	1° varus - 2° valgus
Extension Deficit	7.5° (5.58-30)	Full extension in all patients

DISCUSSION

Customized, patient-specific instruments and implants accurately restore neutral mechanical alignment as measured by intra-operative computer navigation. The patient-specific instruments aligned all patients in this cohort to within $\pm 2^\circ$ of neutral. It is well documented that the restoration of a neutral mechanical axis within $\pm 3^\circ$ is important to achieve long-term survivorship.¹

REFERENCES

1. Ritter MA, Faris PM, Keating EM, et al. Postoperative alignment of TKA. Its affect on survival. *Clin Orthop Relat Res* 1994;299;153
2. Green GV, Berend KR, Berend ME, et al. The effects of varus tibial alignment on proximal tibial surface strain in total knee arthroplasty; the posteromedial hot spot. *J Arthroplasty* 2002; 17;1033
3. Ng VY, DeClaire JH, Berend KR, et al. Improved accuracy of alignment with patientspecificpositioning guides compared with manual instrumentation in TKA. *ClinOrthop Relat Res* 2012;470:99.
4. Nunley RM, Ellison BS, Ruh EL, et al. Are patient-specific cutting blocks costeffective for total knee arthroplasty? *Clin Orthop Relat Res* 2012;470:889.
5. Nunley RM, Ellison BS, Zhu J, et al. Do patient-specific guides improve coronal alignment in total knee arthroplasty? *Clin Orthop Relat Res* 2012;470:895.
6. Lustig S, Scholes CJ, Oussedik SI, et al. Unsatisfactory Accuracy as Determined by Computer Navigation of VISIONAIRE Patient-Specific Instrumentation for Total Knee Arthroplasty. *J Arthroplasty* 2013(28), 469-473.